



Flu Vaccine Clinical Validation Panel

(Catalog No. FLU6230)

For Investigational Research Use Only

Application Notes: The **Influenza Virus vaccine** produced by Lederle Labs and Parke-Davis (FLU-Immune and Fluo-gen) has been implicated in producing false positive results for recipients in a number of infectious disease immunoassays due to what has been characterized as the presence of a “sticky” IgM antibody. The vaccine is a purified surface antigen product prepared from chicken egg extraembryonic fluid. Following administration of the vaccine, recipients may experience slight fever, sore throat, or fatigue. This 13 member, individual subject serum Panel consists of *26 total specimens* collected prior to - *and approximately 2 weeks following* administration of the vaccine.

Panel Member	Age	Sex	Eth	Matrix	Pre Flu Collection Date (A)	Vaccination Date	Post Flu
							Collection Date (B)
FLU6230-01	48	M	Cau	Serum	10/10/95	10/10/95	10/24/95
FLU6230-02	42	M	Cau	Serum	10/10/95	10/10/95	10/23/95
FLU6230-03	43	M	Cau	Serum	10/11/95	10/11/95	10/25/95
FLU6230-04	34	M	Cau	Serum	10/11/95	10/11/95	10/27/95
FLU6230-05	18	M	Cau	Serum	10/10/95	10/10/95	10/20/95
FLU6230-06	32	M	Cau	Serum	10/11/95	10/11/95	10/25/95
FLU6230-07	62	M	Cau	Serum	10/12/95	10/12/95	10/30/95
FLU6230-08	30	M	Blk	Serum	10/17/95	10/17/95	11/01/95
FLU6230-09	45	F	Blk	Serum	10/19/95	10/19/95	11/01/95
FLU6230-10	39	M	Blk	Serum	10/19/95	10/19/95	11/02/95
FLU6230-11	57	M	Cau	Serum	10/22/95	10/22/95	11/05/95
FLU6230-12	50	M	Cau	Serum	10/18/95	10/18/95	11/01/95
FLU6230-13	20	M	Cau	Serum	10/16/95	10/16/95	10/30/95

All sera tested and found to be nonreactive for HBsAg, HCV, and HIV 1/2