

ZeptoMetrix®

NATtrol™ Respiratory Verification Panel Mini

Catalog Number:
NATRMP-BIO

PRODUCT DESCRIPTION:

NATtrol™ Respiratory Verification Panel Mini* (qualitative) is formulated with purified, intact viral particles that have been chemically modified to render them non-infectious and refrigerator stable. NATRMP-BIO contains 6 x 0.6 mL vials of viral NATtrol™ and 2 x 1.8 mL vials of negative (matrix only) as listed in Table 1. The panel members are supplied in a proprietary matrix.

*Pat.: <http://www.zeptometrix.com/patent-information/>

INTENDED USE:

NATtrol™ Respiratory Verification Panel Mini is designed to evaluate the performance of nucleic acid tests for determination of the presence of viral nucleic acids (from organisms listed in Table 1). NATtrol™ Respiratory Verification Panel Mini can also be used for validation of clinical assays, development of diagnostic tests and training of laboratory personnel.

WARNINGS AND PRECAUTIONS:

- NATtrol™ inactivation was carried out on microorganism stocks used to formulate the panel members. The inactivation was verified in a standard microbiological growth protocol.
- This panel contains inactivated microorganisms and materials of human and animal origin. Safe practices suggest that the controls be considered potentially infectious and to use Universal Precautions when handling.
- Refer to CDC guidelines and local regulations for handling and disposal.
- The matrix used in the manufacture of this product is treated with 0.09% sodium azide. It was manufactured from Human Serum Albumin that have been tested and found to be non-reactive at the donor level for HIV-1/HIV-2 Antibody, HBsAg and HCV Antibody by FDA licensed donor screening test methods. All materials are also tested for HIV-1 and HCV by FDA approved Nucleic Acid Test (NAT) methods.
- Heat inactivated Fetal Bovine Serum used in the manufacture of this product meet applicable USDA requirements for abattoir sourced animals, traceability and country of origin. The materials were collected at USDA licensed establishments or legally imported from countries recognized by the USDA as negligible or controlled for risk for Bovine Spongiform Encephalopathy (BSE) and other exotic disease agents. Donor animals were inspected ante and post mortem at the abattoir as required by the USDA.
- Do not use past the expiration date on the label.
- To avoid cross-contamination, use separate pipette tips for all materials.

RECOMMENDED STORAGE:

- NATtrol™ Respiratory Verification Panel Mini should be stored at 2-8°C.

INSTRUCTIONS FOR USE:

- Mix tube vigorously for at least 5 secs.
- Process according to manufacturer's instructions for sample to result assays.
- Extract nucleic acid prior to use in downstream assays that are not sample to result.

LIMITATION:

- FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES**
- Quality control materials should be used in accordance with local, state, federal, and accreditation requirements.
- This product is not intended to replace the manufacturer's controls provided with the assay.

EXPECTED RESULTS:

- Each laboratory must evaluate the product and establish their own acceptance criteria.
- This panel has been tested with the BIOFIRE® SPOTFIRE® Respiratory Panel Mini Assay on the BIOFIRE® SPOTFIRE® System and provides all expected results for the panel members listed in Table 1.








TABLE 1: PANEL MEMBERS

Panel Member	Strain
Influenza A H1N1pdm	A/NY/02/09 ¹
Influenza AH3	A/Brisbane/10/07
Influenza B	B/Florida/02/06
Rhinovirus 1A	N/A
RSV A	N/A
SARS-CoV-2	USA-WA1/2020 ²
Negative	N/A

¹ Please note that although similar in nomenclature, **this is a 2009 H1N1 pandemic Influenza strain** and does NOT correlate with the seasonal 2009 Influenza strains found in the Fludb.org database. For reference, the NCBI Taxon IDs for the seasonal Influenza strains listed in the Fludb.org database are: A/New York/01/2009 (H1N1) - 666252; B/New York/01/2009 - 664512; A/New York/02/2009 (H1N1) - 666298; and A/New York/03/2009 (H3N2) - 659637.

² This reagent was deposited by the Centers for Disease Control and Prevention and obtained through BEI Resources, NIAID, NIH: SARS-Related Coronavirus 2, Isolate USA-WA1/2020, NR-52281.

PINATRMP-BIO
Revision: 01
Effective Date: 05/24/2023

	Catalog Number		Temperature Limitation
	Batch Code		Expiration Date
	For Research Use Only		Biological Risk
	Manufacturer		

PCA# 23-190
Page 1 of 1