

## Certificate of Analysis

**Product Description:** HIV-1 p24 Antigen ELISA 2.0  
**Catalog Number:** 0801002  
**Lot Number:** 327770  
**Expiration Date:** 2022/10/21  
**Storage Temperature:** 2-8°C

### Use Universal Precautions!

<u>Analysis</u>	<u>Specification</u>	<u>Lot Results</u>
<b><i>Optical Density (O.D.) of Antigen Standards:</i></b>		
125 pg/ml	O.D. $\geq$ 1.600	Conforms
3.9 pg/ml	O.D. $\geq$ O.D. of 0.0 pg/ml + 0.030	Conforms
0.0pg/ml	O.D. $\leq$ 0.100	Conforms
<b><i>Labeling:</i></b>	Consistent with Master Production Record	Conforms
<b><i>Storage Temperature:</i></b>	2-8 °C	Conforms
<b><i>Batch Production Record:</i></b>	Complete and Accurate	Conforms

**CAUTION: NOT FOR USE IN HUMANS. FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.**  
 These products are intended for research, product development, quality assurance or further manufacturing use only. These products are NOT intended for use in the manufacturing or processing of injectable products subject to licensure under section 351 of the Public Health Service Act or for any other product intended for administration to humans.

**Statement of Conformance:**

The results reported in this Certificate of Analysis have been certified by ZeptoMetrix’s Quality Assurance department as accurately reflecting the data contained in the batch production and testing records.