

Certificate of Analysis

Product Description: HIV-1 p24 Antigen ELISA 2.0
Catalog Number: 0801002
Lot Number: 329203
Expiration Date: 2023/02/24
Storage Temperature: 2-8°C

Analysis	Specification	Result
ELISA: Optical Density (O.D.) of Antigen Standards		
125 pg/mL	O.D. \geq 1.600	Conforms
3.9 pg/mL	O.D. \geq O.D. of 0.0 pg/ml + 0.030	Conforms
0.0 pg/mL	O.D. \leq 0.100	Conforms

FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

Statement of Conformance:

The results reported in this Certificate of Analysis comply with the requirements of ZeptoMetrix Quality Management System and specifications described in the Master Production Record.