ZeptoMetrix®

This product was manufactured in a facility which has a Quality Management System that is ISO 13485 certified.

Certificate of Analysis

Product Description: **NATtrol™ Flu/RSV/SARS-CoV-2 Negative Control**

Virus: **Coxsackievirus Type A9**

Catalog Number: **NATCV9-6C**

Lot Number: **328443**

Expiration Date: **2023-03-18**

Storage Temperature: **2-8°C**

Safety Precautions: **Use Universal Precautions.**

Test	Specification		Result
Nucleic acid detection by PCR	Influenza A	Not Detected	Pass
	Influenza B	Not Detected	Pass
	RSV B	Not Detected	Pass
	SARS-CoV-2	Not Detected	Pass
Inactivation verification*	Inactive		Pass

*Endpoint dilution assays were used to evaluate inactivation of the NATtrol™ Coxsackievirus master stock used in the manufacture of this product. Samples were inoculated into appropriate cells and incubated for 7 days at 37°C. The inoculated cell cultures were observed for signs of cytopathic effects (CPE). All inactivated samples must demonstrate no CPE to verify inactivation.

FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

Statement of Conformance:
The results reported in this Certificate of Analysis comply with the requirements of ZeptoMetrix's Quality Management System and specifications described in the Master Production Record.

ZeptoMetrix LLC • 878 Main Street, Buffalo, NY 14202 USA • 800-274-5487 • www.zeptometrix.com