

## Certificate of Analysis

**Product Description:** NATtrol™ Flu/RSV/SARS-CoV-2 Negative Control  
**Virus:** Coxsackievirus Type A9  
**Catalog Number:** NATCV9-6C  
**Lot Number:** 330490  
**Expiration Date:** 2023-12-07  
**Storage Temperature:** 2-8°C  
**Safety Precautions:** Use Universal Precautions.

Test	Specification		Result
<b>Nucleic acid detection by PCR</b>	Influenza A	Not Detected	Pass
	Influenza B	Not Detected	Pass
	RSV B	Not Detected	Pass
	SARS-CoV-2	Not Detected	Pass
<b>Inactivation verification*</b>	Inactive		Pass

\*Endpoint dilution assays were used to evaluate inactivation of the NATtrol™ Coxsackievirus master stock used in the manufacture of this product. Samples were inoculated into appropriate cells and incubated for 7 days at 37°C. The inoculated cell cultures were observed for signs of cytopathic effects (CPE). All inactivated samples must demonstrate no CPE to verify inactivation.

**FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.**

**Statement of Conformance:**

The results reported in this Certificate of Analysis comply with the requirements of ZeptoMetrix’s Quality Management System and specifications described in the Master Production Record.